

An experimental approach to study the red blood cell dynamics in a capillary tube by biospeckle laser

Martin A. Toderi^{a,b}, Bibiana D. Riquelme^{a,b}, Gustavo E. Galizzi^{a,c,*}

^a*Instituto de Física Rosario (CONICET-UNR), Bv. 27 de Febrero 210 bis, Rosario, Argentina.*

^b*Facultad de Cs. Bioquímicas y Farmacéuticas, UNR, Suipacha 531, Rosario, Argentina.*

^c*Facultad de Cs. Exactas, Ingeniería y Agrimensura, UNR, Pellegrini 250, Rosario, Argentina.*

Abstract

In this study we investigated the human red blood cell (RBC) dynamics by means of biospeckle laser analysis. Blood samples from healthy donors were introduced in a 0.8 mm internal diameter capillary tube, and illuminated with a He-Ne laser in order to obtain the biospeckle pattern from both side and forward scattered light. Experiments were carried out for different concentrations of red blood cells in plasma, from 25% to 50%. Biospeckle parameters such as Correlation Coefficient and Inertia Moment, were calculated for different frequency bandwidths in order to assess their sensitivity and versatility. A filter based on the Discrete Wavelet Transform was used to decompose the registered sample activity. A relation between Inertia Moment and the RBCs to plasma volume ratio was observed. The Correlation Index that measures the level of correlation of biospeckle images was defined and analyzed. This work inquires in a technique that is suitable for the development of novel non-invasive optical tools for clinical diagnosis in vascular pathologies.

Keywords: red blood cell, biospeckle laser, wavelet filtering, cell interaction

*Corresponding author

Email address: galizzi@ifir-conicet.gov.ar (Gustavo E. Galizzi)

1. Introduction

Human red blood cell (RBC) interaction dynamics has a wide spectrum of interest, involving research on RBC mechanical properties [1, 2, 3], theoretical approaches [4, 5, 6, 7], optical and image processing techniques [8, 9, 10, 11] and immediate application in Biomedicine [12, 13]. Consequently, it is of great importance to develop techniques as well as equipment which allow the evaluation and characterization of blood dynamics due to its implications in vascular pathologies such as diabetes and hypertension [14, 15, 16]. In particular, one of the main goals of these investigations is to contribute to the development of non-invasive techniques in order to early detect and monitor vascular diseases.

In vivo, RBC adhesion phenomena have a determinant role in blood flow and vascular resistance. One of the main consequences of the RBC interaction in stasis, *i.e.* without shear stress, is the formation of structures that normally consist of face-to-face linear assemblies called *rouleaux* [17]. Also, this phenomenon occurs simultaneously with the sedimentation of the structures due to gravity [18]. RBC dynamics can be analyzed by several methods and techniques [19, 20, 21, 22, 23, 24] involving different sample volumes and previous preparation steps.

In the last few years, coherent light-based techniques have aroused attention due to their neglectable perturbation in the studied samples [25, 26]. When a scattering medium, *e.g.* a red blood cell suspension, is illuminated with a coherent light source, an interference pattern is observed at the image plane as shown in Fig. 1. This interferometric phenomenon carries information of the morphology of the illuminated sample. The final optical field is presented as a distribution of *speckles*, and when the distribution is time dependent, the phenomenon is called *dynamic speckle*. Moreover, if this variation in time is the consequence of a biological process, it is presented as *biospeckle laser* (BSL) [27], which has been observed as a sensitive way to study the dynamic behavior of blood [28, 29, 30, 31]. Processes such as cell interaction, cytoplasmic fluctuations and biomechanical reactions related to water activity can be responsible for the




Figure 1: Typical speckle pattern obtained from a human blood sample.

variation of intensity of each speckle pattern.

In this work, BSL analysis was applied to study the dynamic behavior of different concentrations of RBCs inside a capillary tube, as a first approximation to the geometry of blood vessels. Determinations were performed obtaining
35 speckle patterns created by side and forward light scattering (SSC and FSC). In this way, we took into account that the interference phenomena develop differently due to the RBC motion and the different points of observation involved. In order to better identify the cell behavior, we applied a Discrete Wavelet Transform (DWT), which makes it possible to represent a signal through a reduced
40 number of factors [32]. The fact that BSL is a non-destructive technique, together with the relatively simple experimental setup, could make it a relevant tool for biological applications.

2. Materials and Methods

2.1. Blood Sample Preparation

45 Whole blood samples were obtained from healthy donors ($n=5$) by venous puncture and anticoagulated with EDTA. Donor age range was from 25 to 30 years old (two women and three men), did not have any pre-existing health problems (e.g. cardiovascular disease, respiratory disease, endocrine disease,

haematological disorders or neoplasm etc.) and had normal hepatic, and renal
50 function, coagulation and complete blood count. The samples were used within
4 hours after their extraction, according to the recommendations of Ref. [33].
RBCs were obtained from whole blood centrifuged for 5 minutes at 2000 rpm
(Paralwall Model PWL 12T). Buffy coat was discarded and plasma was sepa-
rated and reserved for further use. RBCs were washed two times with saline
55 solution. Samples were prepared with different ratios of RBCs to plasma vol-
ume (from 25% to 50% with steps of 5%) mixing the washed RBCs with the
autologous plasma. The RBCs concentration was checked by a Microhematocrit
centrifuge (Rolco CH 24).

Each prepared blood sample was introduced in a 0.8 mm internal diameter
60 capillary tube by means of a syringe just before conducting the experiment.
RBCs interaction phenomenon occurrence was verified using a previously devel-
oped optical chip aggregometer [8], and the integrity of the cells was confirmed
by optical microscopy.

2.2. Experimental Setup

65 The experimental setup is shown in Fig. 2. The light of a He-Ne laser
($\lambda=632.8$ nm, 60 mW, Melles Griot) passed throughout an intensity dimmer
and a 10x microscope lens (LE), which was used to expand the laser beam in
order to uniformly illuminate the sample. Microscope lenses LS and LF, both
4x magnification, captured side and forward scattered beams. Two diaphragms
70 were positioned after the lenses in order to control the biospeckle grain size.
The sample was tightly placed in the optical path of both lenses slightly out of
focus. Mirrors M1, M2 and M3 permitted the observation of the 90° scattered
beam that gives rise to the SSC biospeckle image. When M3 was removed, the
FSC biospeckle image was obtained. Finally, experimental data were recorded
75 by a video camera (Dalsa model CA-D6, 256 by 256 pixels, $10 \mu\text{m}$ per pixel)
externally driven by a frame grabber (Coreco Imaging PC-DIG) and a CPU.

All experiments were conducted at controlled room temperature of 24°C .
Setup was mounted on a Melles Griot optical table to avoid any vibrational




Figure 2: Schematic representation of the setup from the lateral side and showing a frontal view of the capillary tube containing the blood sample.

interference.

80 2.3. Data Acquisition and Analysis

The laser was turned on and left 10 minutes for beam stabilization. Camera and software were initialized and set to a frame rate of 8 frames per second. A 0.8 mm internal diameter capillary tube was previously filled with approximately 30 μL of blood and secured with syringes on both sides. After tightly positioning
85 the sample, RBCs were slowly disaggregated as proposed in Ref. [34]. The recording of the biospeckle signal started 10 seconds after the disaggregation process ended, and the sample was left non perturbed throughout the entire recording process and 1600 uncompressed images were registered corresponding to a 200 s time lapse. By removing or repositioning mirror M3, two stacks
90 of images were recorded per sample corresponding to FSC biospeckle and SSC biospeckle. The procedure was reproduced for each studied sample.

Data were analyzed by custom software as follows. Firstly, a 3D matrix containing the pixel intensity map for each frame was created. In order to better identify the information corresponding to RBC dynamic behavior, a quadrature
95 mirror filter based on the Discrete Wavelet Transform (DWT) was applied to

the recorded data. After signal decomposition, low frequencies and high frequencies were separated having the cut-off at 1/4 of the sampling frequency. Subsequently, the inverse process was performed but neglecting the *low coefficients* and the *high coefficients* respectively for each reconstruction. The final
100 result consists of two 3D matrix containing the information from 0 to f/4 Hz and f/4 to f/2 Hz respectively, being f the sampling frequency.

2.4. Biospeckle parameters

Computational processing algorithms were applied to the original biospeckle signal and the filtered ones. The biospeckle parameters Correlation Coefficient
105 (CC) and Inertia Moment (IM) were calculated in order to asses the RBC interaction phenomena.

The Correlation Coefficient $CC(k)$ between the first image and the subsequent ones is defined as [35]

$$CC(k) = \frac{\langle I(0)I(k) \rangle - \langle I(0) \rangle \langle I(k) \rangle}{[(\langle I^2(0) \rangle - \langle I(0) \rangle^2)(\langle I^2(k) \rangle - \langle I(k) \rangle^2)]^{1/2}}, \quad (1)$$

where k is the frame number that assumes values from $0, 1, \dots, (K - 1)$ being
110 K the total number of images of the sequence recorded during the experiment. $I(k)$ is the pixel intensity map, a 2D matrix composed by the gray levels of each pixel corresponding to frame k , and $\langle \rangle$ denotes the mean value operator, *i. e.* the mathematical average of the matrix elements. This coefficient provides a quantitative comparison of the similarity of each biospeckle pattern $I(k)$ of the
115 temporal series with respect to the first one $I(0)$ chosen as the reference state. Moreover, the plot of the $CC(k)$ as a function of the frame number k gives a quantitative measurement of the dynamics of the process under analysis [35]. This index is related to the motion of the scatterers, in particular the RBCs, whose dynamics is expected to be conditioned by the number of cells present in
120 the sample.

The computation of IM requires the previous calculation of the Time History Speckle Pattern (THSP) [36] and subsequently the Co-occurrence Matrix (COM) [37]. The THSP shows the time evolution of a pixel intensity. In this




Figure 3: Gaussian distribution for 256 randomly chosen points around the center of the image (white circles), used to construct the THSP.

work, analyzed points were randomly selected based on a Gaussian distribution
 125 around the central pixel of the image (see Fig. 3) for each calculation, giving a
 better statistical approach. Figure 4 shows a THSP used to calculate the corre-
 sponding COM. The COM evaluates the dispersion of consecutive pixels in the
 THSP, representing a transition histogram of intensity. The element $COM(i, j)$
 is the number of occurrences of a certain intensity value i that is immediately
 130 followed by an intensity value j . Points are spreaded along the principal diag-
 onal of the matrix as seen in Fig. 5, indicating the *activity* of the sample. The
 aforementioned activity would be attributed to the Brownian motion of single
 cells, the formation of linear structures and the sedimentation process. Briefly,
 the higher the activity is, the wider the spread around the principal diagonal
 135 results.

The spread around the principal diagonal can be quantified by IM, which is
 defined as

$$IM = \sum_i \sum_j \frac{COM(i, j)}{Norm} |i - j|^2, \quad (2)$$

where i and j are the matrix coordinates and $Norm$ is the normalization of the
 co-occurrence matrix [38].




Figure 4: Time History Speckle Pattern (THSP) constructed from a Gaussian distribution of points for a 50% RBCs to plasma volume sample dynamics. Each row represents the time evolution of the intensity of the corresponding pixel.




Figure 5: Co-occurrence Matrix (COM) calculated from the THSP presented in Fig.4.

140 **3. Results**

RBC interactions were studied by BSL analysis in samples of different ratios of RBCs to plasma volume. Variations of the parameters were observed and their sensitivity was evaluated obtaining the following results.

3.1. Correlation Coefficient

145 The CC was calculated for data obtained by DWT filtering. Determinations were focused on the images obtained by FSC biospeckle because the SSC biospeckle data showed loss of correlation in the first 0.5 s for every sample, with the CC rapidly dropping to a minimum value. The CC was calculated setting the first reference frame after 60 s, 120 s and 180 s to study the evolution in
150 different moments of the RBC interaction process. Fast loss of correlation was observed for the high frequency band regardless the ratio of RBCs to plasma volume and the initial time considered for the CC. On the other hand, the CC showed variations in time for the low frequency band, implying different decorrelation evolution (see Fig 6 (a)). In addition, variations were observed with
155 respect to the ratio of RBCs to plasma volume. In every studied sample, CC values rapidly dropped to a minimum when using the first acquired image as the reference state. Therefore, these particular results were not reported for the sake of clarity. Considering the CC decay as an exponential function, it is possible to linearize the CC dependency by applying the natural logarithm.
160 The natural logarithm of the CC was linearly fitted (see Fig. 6 (b)), and the slope obtained was defined as the Correlation Index (CI). The Coefficient of Determination R-Squared for every linear fitting took values from 0.95 to 0.99. Figure 7 shows the mean values of the CI for the experiments performed by forward scattering, which shows that the CI stabilizes in approximately -2 at
165 40% of RCBs to plasma volume.

3.2. Inertia Moment

IM was calculated for different quantities of analyzed points, starting from 500 up to 10000 points with steps of 500 points as previously described. The




Figure 6: CC for FSC biospeckle data after DWT low pass filtering (a) and linearization of the CC (b) corresponding to 50% of RBCs to plasma volume, inset graphic shows the CI for every initial time.




Figure 7: CIs calculated from FSC biospeckle data corresponding to different ratios of RBCs to plasma volume. The values correspond to the mean obtained from five samples for each ratio of RBCs to plasma volume.

computing time was less than 2 minutes for the whole process per concentration
170 per sample. Calculations were performed using the original signal. The IM
values obtained from iterations with more than 4000 points of analysis showed
insignificant variations for every RBC concentration. Consequently, this number
of analyzed points was proposed for future calculations. Figure 8 (a) shows the
IM value variation versus the number of analyzed points for the particular case
175 of a 35% of RBCs to plasma volume sample. It can be noticed that after 4000
analyzed points the IM values tend to their mean value.

As the ratio of RBCs to plasma volume increases, light intensity is dimin-
ished due to a higher number of cells present in the sample. To address this
issue, light intensity was adjusted to achieve the best contrast conditions regard-
180 ing the compatibility speed between the camera and the phenomenon. Figure 8
(b) shows the mean light intensity versus time for four different ratios of RBCs
to plasma volume. It can be noticed that the mean light intensity is approxi-
mately the same for every experiment and remains roughly constant with time
throughout each measurement.

185 Figure 8 (c) and (d) depict the dependence of the IM on the RBC concen-
tration present in each test. In the case of FSC biospeckle, the IM values are
lower for higher number of cells but the tendency is reversed for SSC biospeckle.
The reported IM values correspond to the mean obtained from five samples for
each ratio of RBCs to plasma volume.




Figure 8: (a) IM values for several iterations of a random Gaussian distribution of selected points for a 35% RBCs to plasma volume concentration sample. (b) Mean light intensity during the measurement and for different RBCs to plasma volume ratios. IM values for different RBCs concentrations obtained for (c) FSC biospeckle and (d) SSC biospeckle.

190 4. Discussion

In this work a DWT based filter was applied to better distinguish RBC interactions, where blood samples were introduced in a capillary tube and illuminated with a He-Ne laser to obtain biospeckle images. Two arrangements were applied, SSC and FSC, obtaining the corresponding correlation coefficients
 195 for the high and low frequency bandwidth data.

This technique is based on collecting the random scattered light from the sample. Therefore, no geometrical reference is associated with the observed speckle pattern. The recorded information is the result of the contribution of all the particles that constitute the analyzed volume. The speckle grains
 200 are consequence of constructive and destructive interference of the scattered light, thus there is no actual visualization of the RBCs. Consequently, this technique does not distinguish inhomogeneities in the particle density inside the capillar related to geometrical aspects. In addition, eventual fluctuations of

the hematocrit inside the tube would be masked by the global behavior of the
205 blood sample.

Results for the high frequency data showed the same behavior disregarding
the blood sample analyzed. The suspension medium, *i.e.* the plasma, is the
variable which stays the same for every test. On the other hand, the CC evolu-
210 tion calculated from the low pass filtering shows variations with respect to the
ratio of RBCs to plasma volume, meaning differences in the dynamics of the
cells. These results suggest that the data from the high frequency bandwidth
mainly contain information about the suspension medium fluctuations as simi-
larly stated in Refs. [39, 40]. Even though the initial and acquisition conditions
were the same for every test, the CC calculated from SSC biospeckle rapidly
215 drops to zero, this not being the case of the calculations for FSC biospeckle.
Consequently, the plasma fluctuations and RBC interaction could have a greater
influence on SSC biospeckle.

It was observed that IM values follow an inverted tendency for data collected
from each experimental setup. IM values obtained by SSC biospeckle increase
220 as the ratio of RBCs to plasma volume increases, but those obtained by FSC
present an inverse behavior. It is important to take into account that the con-
sequences of RBC interaction and plasma characteristics, such as the formation
of 3D structure aggregates known as *rouleaux*, could have a major influence on
the sedimentation process of the scatterers. Through the experiments, sedimen-
225 tation dynamics is observed from different points of view, which may possibly
influence the final result of the IM values, *e. g.* the aforementioned inverted ten-
dency, adding the fact that the volume of RBCs tweaks the amount of activity
present in the sample.

The variations of the IM with respect to the RBCs to plasma volume encour-
230 ages the use of this index as an hematocrit identifier. For each test, the light
intensity was set so as to have the best biospeckle image quality, and therefore
the mean intensity is not determined by each specific sample characteristics.
This fact would make it possible to use the obtained IM vs. hematocrit curves
without the requirement of initial calibration while preserving the biospeckle

235 image quality, as long as the experimental setup is conserved.

5. Conclusion

In the present work we studied the human RBC dynamics by means of biospeckle laser analysis, emulating the geometry of a vessel with a capillary tube. It was possible to obtain two kinds of biospeckle images from one experimental setup corresponding to the SSC and FSC by simply repositioning one mirror. It is presumable that the RBC interaction activity information is mainly contained in the low frequency bandwidth determined by DWT filtering of the data obtained by forward scattering biospeckle. The IM parameter could be used as a RBCs to plasma volume identifier.

245 It should be noticed that no alterations on the RBCs or the suspension medium were performed. Therefore, further studies would involve the alteration of the aggregation properties of the RBCs and the protein content of the plasma, both determinant factors in blood viscosity and microcirculation. The simplicity of the experimental arrangement together with the low-cost components required and its potential non-invasive attribute makes the biospeckle method stand out among the coherent light-based techniques. These characteristics make it suitable for assessing erythrocyte dynamics to improve diagnosis and further treatment of vascular pathologies, in particular because of the statistical nature and time dependency of the results. Short term objectives concern 250 evaluating additional biospeckle descriptors in order to obtain as much information as possible from one measurement. This mainly relies on the biospeckle image quality.

These investigations represent the first steps towards the achievement of innovative optical techniques related to blood characterization and diagnosis.

260 Acknowledgments

Authors thank Dr. Mabel D'Arrigo director of the Clinical Analytical Chemistry Laboratory of the FCByF (UNR) and Dr. Anala Alet for the collaboration

and direction in the preparation of biological samples. This work was partially supported by grant PI BIO400 from UNR and grant PIP 11220150100607 from
265 CONICET, Argentina.

References

- [1] E. Evans, R. Hochmuth, Membrane viscoelasticity, *Biophysical Journal* 16 (1) (1976) 1 – 11. doi:10.1016/S0006-3495(76)85658-5.
- [2] B. Riquelme, J. Valverde, R. Rasia, Complex viscoelasticity of normal and
270 lectin treated erythrocytes using laser diffractometry, *Biorheology* 35 (4) (1998) 325 – 334. doi:10.1016/S0006-355X(99)80014-6.
- [3] B. D. Riquelme, J. R. Valverde, R. J. Rasia, Determination of the complex viscoelastic parameters of human red blood cells by laser diffractometry, *Proc. SPIE, Optical Diagnostics of Biological Fluids V* 3923 (2000). doi:
275 10.1117/12.387133.
- [4] E. Kaliviotis, M. Yianneskis, Blood viscosity modelling: influence of aggregate network dynamics under transient conditions, *Biorheology* 48 (2) (2011) 127–47.
- [5] S. K. Doddi, P. Bagchi, Three-dimensional computational modeling of
280 multiple deformable cells flowing in microvessels, *Phys. Rev. E* 79 (2009) 046318. doi:10.1103/PhysRevE.79.046318.
- [6] F. Janoschek, F. Toschi, J. Harting, Simplified particulate model for coarse-grained hemodynamics simulations, *Phys. Rev. E* 82 (2010) 056710. doi: 10.1103/PhysRevE.82.056710.
- [7] T. Ye, N. Phan-Thien, C. T. Lim, L. Peng, H. Shi, Hybrid smoothed
285 dissipative particle dynamics and immersed boundary method for simulation of red blood cells in flows, *Phys. Rev. E* 95 (2017) 063314. doi: 10.1103/PhysRevE.95.063314.

- 290 [8] M. A. Toderi, H. Castellini, B. Riquelme, Descriptive parameters of the erythrocyte aggregation phenomenon using a laser transmission optical chip, *Journal of Biomedical Optics* 22 (1) (2017) 017003/1–8. doi:10.1117/1.JBO.22.1.017003.
- [9] G. Popescu, Y. Park, W. Choi, R. R. Dasari, M. S. Feld, K. Badizadegan, Imaging red blood cell dynamics by quantitative phase microscopy, *Blood Cells, Molecules, and Diseases* 41 (1) (2008) 10 – 16. doi:10.1016/j.bcmd.2008.01.010.
- [10] G. Jiao, R. Zhang, Modeling of micropipette aspiration and optical tweezers stretching of erythrocytes with or without malaria parasite, *Theoretical and Applied Mechanics Letters* 3 (3) (2013) 034001. doi:10.1063/2.1303401.
- 300 [11] E. Kaliviotis, J. M. Sherwood, S. Balabani, Partitioning of red blood cell aggregates in bifurcating microscale flows, *Scientific Reports* 7 (2017) 44563. doi:10.1038/srep44563.
- [12] S. Shin, Y. Yang, J.-S. Suh, Measurement of erythrocyte aggregation in a microchip stirring system by light transmission, *Clinical Hemorheology and Microcirculation* 41 (3) (2009) 197–207.
- 305 [13] P. Ponce de Leon, M. Toderi, H. Castellini, B. Riquelme, In vitro alterations of erythrocyte aggregation by action of trichinella spiralis newborn larvae, *Clinical Hemorheology and Microcirculation* 65 (2) (2017) 195–204. doi:10.3233/CH-16158.
- 310 [14] B. Riquelme, M. D’Arrigo, P. Foresto, R. Rasia, Laser diffractometry technique for determination of stationary and dynamics viscoelastic parameters of erythrocyte in vascular pathologies, *Proc. SPIE, Optical Coherence Tomography and Coherence Techniques* 5140 (2003). doi:10.1364/ECBO.2003.5140_229.
- 315 [15] M. Delannoy, A. Fontana, M. D’Arrigo, B. Riquelme, Influence of hyperten-

sion and diabetes mellitus on erythrocyte aggregation using image digital analysis, *Series on Biomechanics* 29 (1) (2015) 5–10.


- [16] Y. I. Cho, M. P. Mooney, D. J. Cho, Hemorheological disorders in diabetes mellitus, *Journal of diabetes science and technology* 2 (6) (2008) 11301138. doi:10.1177/193229680800200622.
- [17] S. Chien, L. A. Sung, Physicochemical basis and clinical implications of red cell aggregation, *Clinical Hemorheology* 7 (1987) 71–91.
- [18] T. Fabry, Mechanism of erythrocyte aggregation and sedimentation, *Blood* 70 (5) (1987) 1572–1576.
- [19] O. K. Baskurt, H. Meiselman, E. Kayar, Measurement of red blood cell aggregation in a plateplate shearing system by analysis of light transmission, *Clinical Hemorheology and Microcirculation* 19 (4) (1998) 307–314.
- [20] R. Rasia, N. de Isla, L. Altube, J. Stoltz, J. Valverde, Determination of adhesive specific energy of erythrocyte agglutination by laser retrodiffusion, *Optics and Lasers in Engineering* 39 (5) (2003) 599 – 607. doi:10.1016/S0143-8166(02)00051-9.
- [21] C. Lacombe, J. Lelivre, Interpretation of rheograms for assessing rbc aggregation and deformability, *Clinical Hemorheology and Microcirculation* 7 (1987) 46–61.
- [22] S. Chen, G. Barshtein, B. Gavish, Y. Mahler, S. Yedgar, Monitoring of red blood cell aggregability in a flow-chamber by computerized image analysis, *Clinical Hemorheology and Microcirculation* 14 (4) (1994) 497–508.
- [23] M. Boynard, J. Lelievre, R. Guillet, Aggregation of red blood cells studied by ultrasound backscattering, *Biorheology* 24 (5) (1987) 451–461.
- [24] R. M. Bauersachs, R. B. Wenby, H. J. Meiselman, Determination of specific red blood-cell aggregation indexes via an automated-system, *Clinical Hemorheology and Microcirculation* 9 (1989) 1–25.


- [25] K. Basak, G. Dey, M. Mahadevappa, M. Mandal, D. Sheet, P. K. Dutta,
345 Learning of speckle statistics for in vivo and noninvasive characterization
of cutaneous wound regions using laser speckle contrast imaging, *Microvas-*
cular Research 107 (2016) 6 – 16. doi:10.1016/j.mvr.2016.04.008.
- [26] T. Kyoden, S. Naruki, S. Akiguchi, H. Ishida, T. Andoh, Y. Takada, N. Mo-
mose, T. Homae, T. Hachiga, In vivo visualization method by absolute
350 blood flow velocity based on speckle and fringe pattern using two-beam
multipoint laser doppler velocimetry, *Journal of Applied Physics* 120 (8)
(2016) 084701. doi:10.1063/1.4961611.
- [27] H. J. Rabal, R. A. Braga Jr. (Eds.), *Dymanic Laser Speckle and Applica-*
tions, CRC Press, 2008.
- [28] Y. Aizu, T. Asakura, Bio-speckle phenomena and their application to the
355 evaluation of blood flow, *Optics & Laser Technology* 23 (4) (1991) 205–219.
doi:10.1016/0030-3992(91)90085-3.
- [29] R. Farraro, O. Omid Fathi, B. Choi, Handheld, point-of-care laser speckle
imaging, *Journal of Biomedical Optics* 21 (2016) 21–21–6. doi:10.1117/
1.JBO.21.9.094001.
- 360 [30] S. M. White, R. Hingorani, R. P. Arora, C. C. Hughes, S. C. George,
B. Choi, Longitudinal in vivo imaging to assess blood flow and oxygenation
in implantable engineered tissues, *Tissue Engineering Part C: Methods*
18 (9) (2012) 697–709. doi:10.1089/ten.tec.2011.0744.
- [31] Y. Piederriere, J. L. Meur, J. Cariou, J. Abgrall, M. Blouch, Parti-
365 cle aggregation monitoring by speckle size measurement; application to
blood platelets aggregation, *Opt. Express* 12 (19) (2004) 4596–4601. doi:
10.1364/OPEX.12.004596.
- [32] M. Vetterli, C. Herley, Wavelets and filter banks: theory and design, *IEEE*
370 *Transactions on Signal Processing* 40 (9) (1992) 2207–2232. doi:10.1109/
78.157221.


- [33] O. K. Baskurt, M. Boynard, G. C. Cokelet, P. Connes, B. M. Cooke, S. Forconi, F. Liao, M. R. Hardeman, F. Jung, H. J. Meiselman, G. Nash, N. Nemeth, B. Neu, B. Sandhagen, S. Shin, G. Thurston, J. L. Wautier, New guidelines for hemorheological laboratory techniques, *Clinical Hemorheology and Microcirculation* 42 (2) (2009) 75–97. doi:10.3233/CH-2009-1202.
- [34] O. K. Baskurt, M. Uyklu, S. Ozdem, H. J. Meiselman, Measurement of red blood cell aggregation in disposable capillary tubes, *Clinical Hemorheology and Microcirculation* 47 (4) (2011) 295–305. doi:10.3233/CH-2011-1411.
- [35] A. Federico, G. H. Kaufmann, G. E. Galizzi, H. Rabal, M. Trivi, R. Arizaga, Simulation of dynamic speckle sequences and its application to the analysis of transient processes, *Optics Communications* 260 (2) (2006) 493–499. doi:10.1016/j.optcom.2005.11.047.
- [36] A. Oulamara, G. Tribillon, J. Duvernoy, Biological activity measurement on botanical specimen surfaces using a temporal decorrelation effect of laser speckle, *Journal of Modern Optics* 36 (2) (1989) 165179. doi:10.1016/S0030-3992(99)00033-X.
- [37] R. Arizaga, M. Trivi, H. Rabal, Speckle time evolution characterization by the co-occurrence matrix analysis, *Optics & Laser Technology* 31 (2) (1999) 163 – 169. doi:10.1016/S0030-3992(99)00033-X.
- [38] R. Cardoso, R. Braga, Enhancement of the robustness on dynamic speckle laser numerical analysis, *Optics and Lasers in Engineering* 63 (2014) 19–24. doi:10.1016/j.optlaseng.2014.06.004.
- [39] R. Braga, G. Horgan, A. Enes, D. Miron, G. Rabelo, J. Barreto Filho, Biological feature isolation by wavelets in biospeckle laser images, *Computers and Electronics in Agriculture* 58 (2) (2007) 123–132. doi:10.1016/j.compag.2007.03.009.


- [40] R. A. Braga, L. Dupuy, M. Pasqual, R. R. Cardoso, Live biospeckle laser imaging of root tissues, *European Biophysics Journal* 38 (5) (2009) 679–686. doi:10.1007/s00249-009-0426-0.


400
























Manuscript ID: OLEN_2019_863

Title: An experimental approach to study the red blood cell dynamics in a capillary tube by biospeckle laser

Authors report no financial interests and no potential conflicts of interest originated from this study.